

# PARTICULATE WEIGHING ROOM REPORT

Vehicle ID: **X6XXX8088 / 045M712**      Test ID: **X6XXX8088\_SC032X020719060601 / 1111011467**  
 Test Req.: **082012190660 - 7**      Location: **CHRYSLER TECH CENTER**  
 Test Type: **SC03(2X)**      Facility: **Test Cell 7**  
 Start Time: **06/06/2019 11:35:59**      End Time: **06/06/2019 12:05:48**  
 Requestor: **REDACTED**      Operator: **REDACTED**  
 Seq. Purpose: **MY16 DS baseline with AEM applied**

## Filter Weights

Phase	Filter Type	Stage	Prim. Wt. (mg)	Sec. Wt. (mg)	Total
	Reference	Pre	176.040405	165.451614	
		Post	176.038528	165.452209	
			-----	-----	-----
		Diff.	-0.001877	+	0.000595
				=	-0.001282

### Test Segment: SC03 (3)

Phase	Filter Type	Stage	Prim. Wt. (mg)	Sec. Wt. (mg)	Total
1	Sample	Pre	168.764114	0.000000	
		Post	168.770126	0.000000	
			-----	-----	-----
		Diff.	0.006012	+	0.000000
				=	0.006012

## Results

### Test Segment: SC03 (3)

Phase	CVS Mass (g)	TUN Mass (g)	Sample Ratio	Mass (mg)
1	222843.671	767.965	290.174	1.745

Segment	Mass Per Dist. (mg/Mi)	Mass Per Dist. (mg/Km)
SC03 (3)	0.487	0.303

**Test Validation:**    Valid:    Invalid:    Retest:    Accept:    NIC: system    Date: 06/06/2019 12:19:30  
 Validator's Comments:

## Conditioning

Parameter	Pre Test			Post Test		
	Min.	Max.	Avg.	Min.	Max.	Avg.
Weigh Room Temperature (degC )	22.0	22.0	22.0	22.0	22.1	22.0
Weigh Room Pressure (kPa )	97.9	97.9	97.9	98.1	98.1	98.1
Weigh Room Dew Point (degC )	9.3	9.6	9.5	9.4	9.7	9.5

## Main Events

Event	Date & Time	Event	Date & Time
Pre-Tare Conditioning Started	06/06 06:41	Pre-Tare Conditioning Ended	06/06 07:53
Post-Test Conditioning Started	06/06 12:12	Post-Test Conditioning Ended	06/06 14:29
Tare Weighing Done By Operator	06/06 07:53	Final Weighing Done By Operator	06/06 14:29
Filters Were Put in Sealed Housing	06/06 07:53	Reference Filter Was Changed	05/06 06:59